

OFFICE OF THE CHIEF DISTRICT MEDICAL AND PUBLIC HEALTH OFFICER: GANJAM
CENTRAL STORE, BRAHMAPUR

Letter No. 15140 /CS/2021

Brahmapur

Dated. 07-10-2021

CORRIGENDUM


In the Tender Call Notice uploaded vide this Office Letter No. 14841/CS/2021, Dt. 30.09.2021 and published in the Odia News Paper The Sambada and English News Paper The Times of India on Dt. 02.10.2021, the Following addendum/amendment are made in the Tender:-

- 1) No need of EMD (Rs. 20,000/-) deposit, as per the Letter No. 8943/ F, FIN-COD-MISC-0007-2019, Dt. 18.03.2021 of Government of Odisha, Finance Department.
- 2) No need for attachment of Original Tender Schedule mentioned in the Check List of Annexure-A, Sl. No. 18.
- 3) 3 Nos. of Items added in the Category –A.
- 4) 2 Nos. of Items added in the Category –B.

The Details of the Tender Call Notice and the addendum/amendment may be downloaded from the District website www.ganjam.nic.in.

All other Terms & Conditions of the published Tender will be remained unchanged. Last date of the submission of Tender is remaining unchanged.

CHIEF DISTRICT MEDICAL &
PUBLIC HEALTH OFFICER
GANJAM, BRAHMAPUR



Category -A

Sl No.	Name of the Item	Specification	Name of the Manufacturing Firm	Unit Rate (In Rs.)	GST (In Rs.)	Total Cost (In Rs.)
1	Advanced Nerve Locator with remote control & mapping pen	Detail Specification enclosed as Annexure-A				
2	Single Window Diagnostic Center					
3	Mother Care Kit					

Category -C

1	1.5 Ton Invertor Split AC	3 Star, Including Stand with 5 Years Warranty				
	2 Ton Invertor Split AC	3 Star, Including Stand with 5 Years Warranty				
2	1.5 Ton Window AC	with Stabilizer with Stand with 5 Years Warranty				

I. Nerve Locator SPECIFICATIONS (Pain management)

SL NO	Name of the Item	SPECIFICATION
1	Advanced Nerve Locator with remote control & mapping pen.	Ergonomic shape and large display
		Instrument type: BF
		Battery : Should run on any 9V (alkaline) batteries
		Power consumption : 6 mA (* mA max)
		Stimulation current : 5 mA max (0-12 k), Stimulation Voltage : 95V
		Stimulation frequency : 1 Hz / 2 Hz , Allowable load impedance : 0 k - 12
		Stimulus duration : 1.0 ms - 0.5 ms - 0.3 ms - 0.1 ms - 0.05 ms
		Digital dial for Precise and tactile current adjustment with convenient tactile feedback.
		Sense function stimulation duration- SENSE (0.10 ms - 0.10 ms - 0.15 ms to 1.0 ms)
		Current measuring accuracy : +/- 0.02 mA
		Impedance measuring range : 1 KOhms - 90 Kohms for target stimulation current > 0.5 mA
		Weight : 250 g
		The enlarged full graphics LC display having good visibility with wide angle of view.
		The following information should be displayed on the stimulation screen at all times:
		Stimulus amplitude in mA (large digits)
		Current range :
		Stimulus duration in ms
		Stimulus frequency in Hz
		Load impedance in kΩ
		Direct access keys: Direct access keys to let conveniently adapt the most important and frequently used settings to the clinical environment. (amplitude range - mA, stimulus duration - ms, stimulus frequency - Hz).
		Convenient menu navigation: Adjust your default settings conveniently by navigating through the menu with the arrow keys on the menu keypad. Just press the power on/off button for less than a second to jump back to the stimulation screen.
		Sequential electrical nerve stimulation (SENSe) - Featuring the option of selecting a frequency of either 1 Hz, 2 Hz or 3 Hz for SENSe and a stimulus duration of 0.05 ms - 0.10 ms - 0.30 ms - 0.50 ms - 1.00 ms.
		It should have following Safety Features:
		Threshold current - To visually and acoustically indicate when the stimulus duration-dependent current threshold goes out-of-tolerance indicating that the needle may be too close to the nerve.
		Visual warning: The full digits of the target current display converted to contoured digits when the current threshold drops below the preset level. The LED flashing yellow (instead of green).
		Acoustic warning: A warning signal should be additionally sounded.
		Impedance Check: The HNS 12 alarm for unacceptable impedance levels that can be caused by a loose electrode cable or dry skin electrode.
		Auto Adjust current feature - the current adjusting automatically when the stimulus duration is changed.
Optional Accessories:		
Remote control: That allows sterile one-handed operation.		
Pen: The Pen for percutaneous nerve mapping to help to pre-assess the puncture site. Also suitable for training sessions.		
Disposable stimulation needles for single shot & continuous technique:		

2. SINGLE WINDOW DIAGNOSTIC CENTER

SL. No.	ITEM	SPECIFICATIONS
1	ESR ANALYSER	1. Principle: Photometric Rheoscope 2. Analytical Range : 1-130 mm/hr 3. Sample Requirements: 100ul Whole Blood (500 ul Dead Volume) 4. Tube Requirements: EDTA 5. Barcode Reader: Internal 6. Power Requirements: 100-240 VAC; 50-60 Hz; 160W
2	Urine Analyser	1. Test parameters: urobilinogen, bilirubin, ketone, blood, protein, nitrite, leukocytes, glucose, specific gravity, pH, ascorbic acid 2. Test wavelength: 525nm, 572nm, 610nm, 660nm 3. Test principle: Photoelectric colorimetry 4. Test throughput: Fast Mode: 120 Test/h, Slow Mode: 60 Test/h 5. Data memory: 1000 patient results & 50 Q.C. results 6. Computer interface: RS-232 port; parallel printer interface 7. Display: 130 x 40 mm LCD 8. Power Supply: 100-240VAC, 50Hz/60Hz 9. Printer: Built-in thermal printer / external printer (optional)
3	Immunoassay Analyzer	1. Technology: Immuno Fluotrescence 2. Test Menu : 25 Parameters 3. 8 Inches LCD touch screem 4. Built in Thermal printer 5. LIS connectivity 6. 30000 Test results storage 7. Results in 3-15 minutes 8. Self calibration
4	Hemoglobin Analyzer	1. Results in 1 Second 2. Measuring Range: 0 - 25.5 g/dL 3. Automatic Self-Check between every measurement 4. Microcuvettes unaffected by Humidity or Temperature & can be used till expiry. 5. Integrated Rechargeable Lithium-ion battery. 6. Factory calibrated against HICN reference method according to ICSH.
5	Blood gas Analyser	1. Results in 1 Minute after sample aspiration. 2. Sample Volume - 140ul 3. Quality Control : 3 or 5 levels QC, External electronic simulator 4. Display : 7 inches Colour LCD display 5. 4200 mAH rechargeable Lithium-ion battery

CONSUMABLES

Sl.No	Name Of The Items	PACK SIZE
SL.No.1	ESR ANALYZER	
	ESR CARTRIDGE	500 Units
SL. NO. 2	URINE ANALYZER	
	10 PARAMETERS	100 Tests / Unit
SL. NO. 3	IMMUNOASSAY ANALYZER	
	Parameter	
	CRP	25 Units
	PCT	25 Units
	Troponin I	25 Units
	cTnI/CK-MB/MYO (3 in 1)	25 Units
	BNP	25 Units
	NT-proBNP	25 Units
	CK-MB	25 Units
	MYO	25 Units
	cTnI/proBNP	25 Units
	AFP	25 Units
	PSA	25 Units
	CEA	25 Units
	HbA1c	25 Units
	MAU	25 Units
	NGAL	25 Units
	b-HCG	25 Units

	Prog	25 Units
	PRL	25 Units
	LH	25 Units
	FSH	25 Units
	T3	25 Units
	T4	25 Units
	TSH	25 Units
	Vit D	25 Units
SL. NO. 6	Hemoglobin Analyzer	
	HEMOGLOBIN CUVETTES	500 Units
SL. NO. 7	Blood Gas Analyzer	
	8 Parameter Cartridge	25 Units
	10 Parameter Cartridge	25 units

3. MOTHER CARE KIT SPECIFICATION

SL.NO.	NAME OF THE ITEMS	SPECIFICATION
1	Night Gown For Mother	<ul style="list-style-type: none"> • Fabrics - Cotton. • Types - Maternity/Nursing Nighty. • Suitable For - Maternity Wear. • Closure - Zipper. • Length - Long. • Other Details - Comfortable Smart and Elegant Feeding Nighty / Nursing Nighty For Everey Lactating Mother:
2	Postnatal Maternity Abdominal Belt / Binder	<ul style="list-style-type: none"> • Made of premium quality breathable material for superior comfort. • Helps in pain relif and correct posture both during pregnancy and post delivery. • Adjustable belt for women of all body shapes and sizes. <p>SPECIFICATION:</p> <ul style="list-style-type: none"> • Type - Post Natal Maternity. • Abdominal Belt - Support to comress the abdominal muscle for waist line reduction and is used as a post operative / post delivery Aid. • Product Dimension - 110 (L) x 23 (B) cm, 121 (L) x 23 (B) cm, 124 (L) x 23 (B) cm. • Size - Small (28" - 32"), Medium (32" - 36"), Large (36" - 40"), XXL (44" - 48"). • Extra porous three panel webbing improves ventilation and comfort of the patient. • Broad hook and loop tape panel offers better holding and size adjustability. • Optimal compression, tones up abdominal muscles following surgery, CS or delivery. • Optimal compression binds surgical incisions of abdomen & chest, improves healing. • Special nylon reeves prevent rolling over of belt.
3	Rechargeable Torch	<ul style="list-style-type: none"> • 6 V 3.8Ah SMF battery. • 6V 0.5A focus bulb. • Backup of app 6 hrs. • Side lamp for spread light.
4	Tooth Brush and Paste	<p>Tooth Brush:</p> <ul style="list-style-type: none"> • Material - Plastic. • Type - Medium. • Tooth Brush Type - Soft. <p>Tooth Paste:</p> <ul style="list-style-type: none"> • Tooth Paste contains the following ingredients - Binders, abrasives, sudsers, humectants, flavors (unique additives), sweetners, fluorides, tooth whiteners, a preservative, and water. • Binder thicken toothpaste. • They prevent separation of the solid and liquid components, especially during storage. • They also effect the speed nand volume of foam production, the rate of flavor release and product dispersal, the appearance of the toothpaste ribbon on the toothbrush, and the rinsibility from the toothbrush. • Some binders are karaya gum, bentonite, sodium alginate, methylcellulose, carrageenan, and magnesium aluminum silicate.

5	Bucket, Mug and Water Bottle	<ul style="list-style-type: none"> • Ltrs - 5 ltrs. • Height - 20 inch. • Top inch - 22. • Diameters Bottom inch - 18. • Material - Commercial / Standard. <p>Mug:</p> <ul style="list-style-type: none"> • Material - Polypropylene. • Size - Frosty 750 - D - 100mm, H - 111mm - 700ml. Frosty 1000 - D - 120mm, H - 130mm - 1000ml. Frosty 1500 - D - 135mm, H - 14mm - 1500ml. <p>Water Bottle:</p> <ul style="list-style-type: none"> • Capacity - 1100ml (Set of 3). • BPA Free. • Japanese Technology. • Leak Proof. • Break Proof. • Crystal Clear. • 100% food Grade. • Hygienic. • Refrigerator safe.
6	Paladai, Tumbler, Bowl & Spoon for Feeding Baby After	<ul style="list-style-type: none"> • Type - Bowl. • Capacity - 10ml. • Suitable for drinking. • Color - Steel. • Material - steel. <p>Tumbler:</p> <ul style="list-style-type: none"> • Material Grade - SS 304. • Capacity - 180 ml. • Color - Silver. • Finish Type - Mirror Polish. • Material - Stainless Steel. <p>Bowl:</p> <ul style="list-style-type: none"> • Upper Length - 12.5 cm. • Lower Length - 7.5 cm. • Deep - 5 cm. <p>Spoon:</p> <ul style="list-style-type: none"> • Size - 13 cm. • Type - Utensil Combo. • Sterilizable - Yes. • Color - Stainless Steel.

NOTE:

- At the time of supply the supplier must submit a test report of the Batches being supplied from approved Laboratory that, the supplied items are of standard quality.